

510(k) Summary of Safety & Effectiveness

Pursuant to CFR 807.92, the following 510(k) Summary is provided:

1. (a) **Submitter Address:** George J. Hattub
MedicSense, USA
291 Hillside Avenue
Somerset, MA 02726
www.medicsense.com
1. (b) **Manufacturer Address:** Biological Signal Processing Ltd.
2a Habarzel Street
Tel-Aviv 69710, Israel

Mfg. Phone: Tel.: 972-3-647 4840

Contact Person: Eran Toledo

Date: August 20, 2007
2. **Device & Classification Name:** Monitor, Physiological, Patient (Without Arrhythmia Detection Or Alarms)
device as a class II device (product code MWI).
HyperQ
3. **Predicate Device:** HyperQ™ System K070624
4. **Description:** The HyperQ™ System is a compact monitor for measuring, processing, storing, and displaying information derived from an electrocardiogram (ECG). The device analyzes and records the high frequency components of the QRS complex of standard ECG.
5. **Intended Use:** ECG

ECG is intended to disclose either normal condition or patterns of arrhythmia, myocardial ischemia, rate abnormalities, or features of prognostic value in the following cases:
Patients with suspected cardiac abnormalities
Populations of patients at an age or period in which a routine baseline evaluation of ECG characteristics is desired.

Stress Testing

Angina pectoris (chest pain) is a clinical syndrome resulting from myocardial ischemia, indicative of reduced blood supply to the cardiac muscle. The electrocardiogram may establish the diagnosis of ischemic heart disease if characteristic changes are present.
Stress testing is the most widely used method to decide whether this chest

pain is related to myocardial ischemia, and thus to coronary artery disease. In stress testing, the contractile capability of the heart muscle is monitored via ECG during patient exercise. Patients exercise by bicycle, treadmill, or other means, while the ECG is monitored continuously. Exercise loads are determined by predefined protocols. The ECG signals as well as the HF-QRS signals are recorded for the resting, exercise, and recovery phase portions of the exercise protocol. The changes in both ECG waveforms are compared to the resting ECG records. In the HyperQ™ stress test, changes in the high frequency of the mid QRS complex, calculated as root-mean-square (RMS) values, are compared to the resting values. Most of the commercial stress test systems control the bicycle or treadmill automatically according to the requirements of the chosen protocol, although this is not essential.

ST segment monitoring is intended as an aid in the evaluation of myocardial ischemia in patients with known or suspected coronary artery disease. The ST segment algorithm has been tested for accuracy of the ST segment data, and a database is used as a tool for performance testing. The significance of the ST segment changes must be determined by a physician.

HyperQ™

The HyperQ™ Software is intended to be used as an aid to stress ECG test by means of analysis of high frequency components present within the central portion of the QRS complex.

The significance of the HF-QRS changes must be determined by a physician.

6. ***Comparison of
Technological
Characteristics:***

With respect to technology and intended use, the Modified HyperQ™ System is substantially equivalent to its predicate device which is the HyperQ™ System. The primary difference is that the modified device provides a reformatted graphic interface unit (GIU) with additional enhancements, features, and user selectable options. Based upon the validation results, BSP believes these differences do not raise additional safety of efficacy concerns.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

SEP 18 2007

MedicSense, USA
c/o Mr. George J. Hattub, RAC and CQE
Senior Staff Consultant
291 Hillside Avenue
Somerset, MA 02726

Re: K072389
HyperQ™ System
Regulation Number: 21 CFR 870.2300
Regulation Name: Cardiac monitor (including cardiometer and rate alarm)
Regulatory Class: Class II (two)
Product Code: MWI
Dated: August 21, 2007
Received: August 24, 2007

Dear Mr. Hattub:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,


for Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K072389

Device Name: HyperQ™ System

Indications For Use: ECG is intended to disclose either normal condition or patterns of arrhythmia, myocardial ischemia, rate abnormalities, or features of prognostic value in the following cases:

- Patients with suspected cardiac abnormalities
- Populations of patients at an age or period in which a routine baseline evaluation of ECG characteristics is desired.

Stress Testing

Angina pectoris (chest pain) is a clinical syndrome resulting from myocardial ischemia, indicative of reduced blood supply to the cardiac muscle. The electrocardiogram may establish the diagnosis of ischemic heart disease if characteristic changes are present.

Stress testing is the most widely used method to decide whether this chest pain is related to myocardial ischemia, and thus to coronary artery disease. In stress testing, the contractile capability of the heart muscle is monitored via ECG during patient exercise. Patients exercise by bicycle, treadmill, or other means, while the ECG is monitored continuously. Exercise loads are determined by predefined protocols. The ECG signals as well as the HF-QRS signals are recorded for the resting, exercise, and recovery phase portions of the exercise protocol. The changes in both ECG waveforms are compared to the resting ECG records. In the HyperQ™ stress test, changes in the high frequency of the mid QRS complex, calculated as root-mean-square (RMS) values, are compared to the resting values.

Most of the commercial stress test systems control the bicycle or treadmill automatically according to the requirements of the chosen protocol, although this is not essential.

ST segment monitoring is intended as an aid in the evaluation of myocardial ischemia in patients with known or suspected coronary artery disease. The ST segment algorithm has been tested for accuracy of the ST segment data, and a database is used as a tool for performance testing.

The significance of the ST segment changes must be determined by a physician.

HyperQ™

The HyperQ™ Software is intended to be used as an aid to the ECG stress test by means of analysis of high frequency components present within the central portion of the QRS complex.

The significance of the HF-QRS changes **must** be determined by a physician.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Ashley Brann for BDZ
(Division Sign-Off)

Division of Cardiovascular Devices

510(k) Number K072389